

SPECIFICATION SHEET

Product Name:	CA19-9 Antigen Calibrator Grade	Catalog #:	L2C007
Description:	Human Cancer Antigen19-9(CA19-9)Antigen , Calibrator Grade		
Source:	ascites, from the human colorectal adenocarcinoma.		
Format:	Calibrator Grade, Liquid		
Concentration:	1000KUnits/ml, (CanAg ELISA)	Lot:	L181201
Buffer:	20Mm PBS, pH 7.4 ± 0.02,		
Preservative:	0.05% Sodium Azide		
Applications:	Specific methodologies have not been tested using this product.		
Storage:	Store at -20°C		
Inactivation:	Not applicable.		
Warnings:	All cultures have tested negative for HBsAg, HCV, HIV-1 and HIV-2. No test guarantees a product to be non-infectious. Therefore, all material should be considered as potentially infectious.		

This product contains sodium azide, which has been classified as Xn (Harmful) in European Directive 67/548/EEC in the concentration range of 0.1 – 1.0 %. When disposing of this reagent through lead or copper plumbing, flush with copious volumes of water to prevent azide build-up in drains.